

[illegible]

- Graduation 0.001"
- With lug back
- Supplied with limit pointers and bezel clamp
- Optional accessory: backs, contact points

Code	Range	Accuracy		Range/rev	Dial reading	Maximum measuring force	Cap
		First 2.5 rev	Overall				
2326-2*	2"	±0.001"	±0.004"	0.1"	0-100	2.5N	cannot lift

* Supplied with manufacturer inspection certificate